

<Sample Information>


Glouannec Alan
 Sample Name : Critical boost (2022240219)
 Sample ID : CBDeau
 Data Filename : Data_20220302_CBDeau_Critical boost (2022240219).lcd
 Method Filename : HighSensitivity.lcm
 Batch Filename : Analysis20220302_142545.lcb
 Vial # : 1-32
 Injection Volume : 5 uL
 Sample Amount : 201 mg
 Extraction Vol. : 20 mL
 Dilution Factor : 10
 Date Acquired : 02/03/2022 15:18:16
 Date Processed : 03/03/2022 16:27:15

Sample Type : Unknown
 Level : 1
 Quantification limit : 0,01%

Acquired by : Glouannec Alan
 Processed by : Glouannec Alan

<Chromatogram>

mV



<Quantitative Results>

Detector A

ID#	Name	Ret. Time	Dry weight %
1	CBDV	2,681	0,05
2	CBDA	3,556	2,98
3	CBGA	3,841	0,16
4	CBG	4,111	0,04
5	CBD	4,308	16,07
6	THCV	--	0,00
7	CBN	--	0,00
8	d9-THC	6,838	0,01
9	d8-THC	--	0,00
10	CBC	7,677	0,00
11	THCA	7,893	0,15

Total THC	0,14	%
Total THC	1,42	mg/g
Total CBD	18,69	%
Total CBD	186,87	mg/g
Cannabinoid	19,46	%
Cannabinoid	194,60	mg/g

Les résultats de cette analyse ne sont valables que pour l'échantillon ayant été fourni au laboratoire.
 Un échantillon sera conservé pour une période de 3 mois après analyse si une quantité suffisante a été fournie.

Validé par le responsable de laboratoire :

Alan Glouannec



Laboratoire d'analyse de la société
TM Nature distribution
99 Rue des Fleurs - 73170 Yenne
Tel: 06 38 05 83 51
E-mai : contact@test-cbd.com
n°Siret : 84841988300030